## **Biomedical Optics EXPRESS**

## Sensitivity of visible range multi-wavelength algorithms for retinal tissue oximetry to acquisition parameters: supplement

CLÉOPHACE AKITEGETSE, 1,\* D JASMINE POIRIER, 1 AND DOMINIC SAUVAGEAU<sup>1,2</sup>

This supplement published with Optica Publishing Group on 26 July 2023 by The Authors under the terms of the Creative Commons Attribution 4.0 License in the format provided by the authors and unedited. Further distribution of this work must maintain attribution to the author(s) and the published article's title, journal citation, and DOI.

Supplement DOI: https://doi.org/10.6084/m9.figshare.23706216

Parent Article DOI: https://doi.org/10.1364/BOE.495721


<sup>&</sup>lt;sup>1</sup>Zilia inc., Québec, QC, G1K 3G5, Canada

<sup>&</sup>lt;sup>2</sup>Chemical and Materials Engineering, University of Alberta, Edmonton, AB, T6G 1H9, Canada


 $<sup>^*</sup>cleo@ziliahealth.com\\$ 

## Sensitivity of visible range multi-wavelength algorithms for retinal tissue oximetry to acquisition parameters: supplemental document


## SUPPLEMENTAL FIGURES




**Fig. S1.** Selection of the spectral range. Scatter plots of  $StO_2$  values calculated by the multi-wavelength method versus expected values, when the narrow spectral band is set to 510-585 nm, 510-580 nm, 515-590 nm and for comparison, the 530-585 nm range which gives the best agreement. Perfect match is indicated as a dashed line.




**Fig. S2.** Selection of the spectral range. Scatter plots of  $StO_2$  values calculated by the multi-wavelength method versus expected values, when the narrow spectral band is set to 515-585 nm, 515-580 nm, 520-590 nm and for comparison, the 530-585 nm range which gives the best agreement. Perfect match is indicated as a dashed line.



**Fig. S3.** Selection of the spectral range. Scatter plots of StO<sub>2</sub> values calculated by the multi-wavelength method versus expected values, when the narrow spectral band is set to 520-585 nm, 520-580 nm, 525-590 nm and for comparison, the 530-585 nm range which gives the best agreement. Perfect match is indicated as a dashed line.



**Fig. S4.** Selection of the spectral range. Scatter plots of  $StO_2$  values calculated by the multi-wavelength method versus expected values, when the narrow spectral band is set to 525-585 nm, 525-580 nm, 530-590 nm and for comparison, the 530-585 nm range which gives the best agreement. Perfect match is indicated as a dashed line.



**Fig. S5.** Selection of the spectral range. Scatter plots of  $StO_2$  values calculated by the multi-wavelength method versus expected values, when the narrow spectral band is set to 530-580 nm, 540-590 nm, 540-585 nm, 540-580 nm and for comparison, the 530-585 nm range which gives the best agreement. Perfect match is indicated as a dashed line.